

# Quality Control Materials for use with BIOFIRE® FILMARRAY® Panels

## Introduction


---


The Clinical Laboratory Improvement Amendments (CLIA), passed in 1988, establishes quality standards for all laboratory testing to ensure the accuracy and reliability of patient test results, regardless of where the test is performed. CLIA regulations require a laboratory to have quality control (QC) procedures to monitor the accuracy and precision of the complete testing process.

This document provides a listing of QC materials that have been tested at BIOFIRE and are compatible with the BIOFIRE Panels indicated. The information in this document is meant to be a guideline and may not be inclusive.

For guidelines to assist your laboratory in developing an individualized quality control plan for BIOFIRE® FILMARRAY® Systems, please refer to the Example Individualized Quality Control Plan (IQCP) Risk Analysis for the following panels:

BIOFIRE® FILMARRAY® Blood Culture Identification (BCID) Panel: FLM1-PRT-0217  
 BIOFIRE® Blood Culture Identification 2 (BCID2) Panel: BFR0000-6645  
 BIOFIRE® FILMARRAY® Gastrointestinal (GI) Panel: FLM1-PRT-0218  
 BIOFIRE® Joint Infection (JI) Panel: BFR0000-9504  
 BIOFIRE® FILMARRAY® Meningitis/Encephalitis (ME) Panel: FLM1-PRT-0219  
 BIOFIRE® FILMARRAY® Pneumonia Panel: FLM2-PRT-0313  
 BIOFIRE® FILMARRAY® Respiratory Panel (RP): BFR0000-8660  
 BIOFIRE® FILMARRAY® Respiratory Panel 2 (RP2): BFR0000-8660  
 BIOFIRE® FILMARRAY® Respiratory Panel 2 *plus* (RP2*plus*): BFR0000-8660  
 BIOFIRE® Respiratory Panel 2.1 (RP2.1): BFR0000-8660  
 BIOFIRE® Respiratory Panel 2.1 *plus* (RP2.1*plus*): BFR000-8660  
 BIOFIRE® Respiratory Panel 2.1-EZ (RP2.1-EZ): BFR000-8660

 **Note:** The QC materials listed are appropriate for use with the panel described and may not be compatible with other BIOFIRE Panels.

 **Note:** The use of routine quality control materials is not appropriate for verification of test method performance (COM.40300). Please refer to BIOFIRE Laboratory Verification Tech Notes for further instruction on verification protocols.



## Quality Control Resources

**Table 1. QC Materials for BIOFIRE BCID Panel**

BIOFIRE BCID Panel		
Vendor	Product Name	Part Number
Maine Molecular Quality Controls, Inc.	FilmArray® BCID Control Panel	M235
Microbiologics	Blood Culture Identification (BCID) Control Panel (Inactivated)	8215 <sup>1</sup>

<sup>1</sup> For in vitro Diagnostic Use; an FDA cleared Class I Unassayed Control

**Table 2. QC Materials for BIOFIRE BCID2 Panel**

BIOFIRE BCID2 Panel		
Vendor	Product Name	Part Number
Maine Molecular Quality Controls, Inc.	FilmArray® BCID2 Control Panel	M416 <sup>2,3</sup>
Microbiologics	Blood Culture Identification Control Panel (43 Targets)	8254 <sup>1,3</sup>
Streck, Inc	Mdx-Chex™ for BCID2	250065 <sup>2,3</sup>

<sup>1</sup> For in vitro Diagnostic Use; an FDA cleared Class I Unassayed Control

<sup>2</sup> For in vitro Diagnostic Use; an FDA cleared class II Assayed Control

<sup>3</sup> CE-IVD marked

**Table 3. QC Materials for BIOFIRE GI Panel**

BIOFIRE GI Panel		
Vendor	Product Name	Part Number
Maine Molecular Quality Controls, Inc.	FilmArray® GI Control Panel	M238 <sup>1</sup>
ZeptoMetrix Corporation	NATtrol™ GI Controls	NATGIC-BIO

<sup>1</sup> CE-IVD marked



**Table 4. QC Materials for BIOFIRE JI Panel**

BIOFIRE JI Panel		
Vendor	Product Name	Part Number
Maine Molecular Quality Controls, Inc.	BioFire JI Control Panel	M420 <sup>1,2</sup>

<sup>1</sup> CE-IVD marked

<sup>2</sup> For in vitro Diagnostic Use; an FDA cleared class II Assayed Control

**Table 5. QC Materials for BIOFIRE ME Panel**

BIOFIRE® ME Panel		
Vendor	Product Name	Part Number
Maine Molecular Quality Controls, Inc.	BioFire® ME Control Panel	M262 <sup>1</sup>
ZeptoMetrix Corporation	NATtrol™ Meningitis/Encephalitis (ME) Controls	NATMEC-BIO

<sup>1</sup> CE-IVD marked

**Table 6. QC Materials for BIOFIRE Pneumonia Panel and Pneumonia *plus* Panel**

BIOFIRE® Pneumonia and Pneumonia <i>plus</i> Panels		
Vendor	Product Name	Part Number
Maine Molecular Quality Controls, Inc.	FilmArray® Pneumonia/ Pneumonia <i>plus</i> Control	M340 <sup>2,3,4</sup>
Microbiologics	Pneumonia (33 Targets) Control Panel	8244 <sup>1</sup>

<sup>1</sup> For in vitro Diagnostic Use; an FDA cleared Class I Unassayed Control

<sup>2</sup> For in vitro Diagnostic Use; an FDA cleared class II Assayed Control

<sup>3</sup> Control contains MERS-CoV2 recombinant material

<sup>4</sup> CE-IVD marked

**Table 7. QC Materials for BIOFIRE RP**

BIOFIRE RP		
Vendor	Product Name	Part Number
Exact Diagnostics	RP Positive Run Control/ RP Negative Run Control	RPPOS <sup>1,2,3/</sup> RPNEG <sup>1,2</sup>
Maine Molecular Quality Controls, Inc.	FilmArray® RP Control Panel	M210
Microbiologics	Respiratory Control Panel (22 Targets)	8247 <sup>1</sup>
ZeptoMetrix Corporation	NATtrol™ RP Multimarker Controls	MDZ001 <sup>2</sup>

<sup>1</sup> For in vitro Diagnostic Use; an FDA cleared Class I Unassayed Control

<sup>2</sup> CE-IVD marked

<sup>3</sup> Control contains MERS-CoV2 recombinant material

**Table 8. QC Materials for BIOFIRE RP2 and RP2plus**

BIOFIRE® Respiratory 2 Panel (RP2) and Respiratory 2 plus Panel (RP2plus)		
Vendor	Product Name	Part Number
Exact Diagnostics	RP Positive Run Control/ RP Negative Run Control	RPPOS <sup>1,3,4/</sup> RPNEG <sup>1,3</sup>
Maine Molecular Quality Controls, Inc	FilmArray® RP2/RP2plus Control Panel	M315 <sup>2,3,4</sup>
Microbiologics	Respiratory (22 Targets) Control Panel	8247 <sup>1</sup>
ZeptoMetrix, LLC	NATtrol™ Respiratory Panel 2.1 (RP2.1) Controls	NATRPC2.1- BIO

<sup>1</sup> For in vitro Diagnostic Use; an FDA cleared Class I Unassayed Control

<sup>2</sup> For in vitro Diagnostic Use; an FDA cleared class II Assayed Control

<sup>3</sup> CE-IVD marked

<sup>4</sup> Control contains MERS-CoV2 recombinant material



**Table 9. QC Materials for BIOFIRE RP2.1, RP2.1*plus*, and RP2.1-EZ**


BIOFIRE® Respiratory 2.1 Panel (RP2.1)*		
Vendor	Product Name	Part Number
Exact Diagnostics	RP Positive Run Control/ RP Negative Run Control	RPPOS <sup>1,3, 4</sup> / RPNEG <sup>1,3</sup>
Maine Molecular Quality Controls, Inc.	BioFire RP2.1/RP2.1 <i>plus</i> Control Panel	M441 <sup>2,3,4</sup>
Microbiologics	Respiratory Control Panel (22 Targets)	8247 <sup>1</sup>
ZeptoMetrix, LLC	NATtrol™ Respiratory Panel 2.1 (RP2.1) Controls	NATRPC2.1- BIO

<sup>1</sup> For in vitro Diagnostic Use; an FDA cleared Class I Unassayed Control

<sup>2</sup> For in vitro Diagnostic Use; an FDA cleared class II Assayed Control

<sup>3</sup> CE-IVD marked

<sup>4</sup> Control contains MERS-CoV2 recombinant material

 **Note:** Quality control options are discussed more thoroughly in *External Control Material Options for BIOFIRE® Respiratory Panel 2.1 (RP2.1)* BFR000-8900.

## Technical Support Contact Information

bioMérieux is dedicated to providing the best customer support available. If you have any questions or concerns about this process, please contact the BIOFIRE Technical Support team for assistance.

BIOFIRE Technical Support  
 Email: [biofiresupport@biomerieux.com](mailto:biofiresupport@biomerieux.com)  
 Phone: +1-801-736-6354, select Option 5

\*All product names, trademarks and registered trademarks are property of their respective owners.